



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

ITRC

38130 Airport Parkway, Unit #1, Willoughby, OH 44094

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Mechanical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Initial Accreditation Date:

May 23, 2011

Issue Date:

April 24, 2017

Expiration Date:

July 31, 2019

Accreditation No.:

66183

Certificate No.:

L17-172

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

ITRC

38130 Airport Parkway, Unit #1, Willoughby, OH 449094
 Contact Name: John Frangowlakis Phone: 440-622-9540

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Master Torque Transducers ^F	15 lbf·ft to 16.67 lbf·ft	0.03 lbf·ft	Dead Weight Testing System Class F Weights and Beams calibrated with results traceable to the SI.
	16.68 lbf·ft to 1 000 lbf·ft	1.7 lbf·ft	
Transducer Tools Including DC/Electric, Torque Wrenches, Torque Analyzers, Air/Pneumatic ^{FO}	0.5 lbf·ft to 4.16 lbf·ft	0.02 lbf·ft	Master Transducers calibrated with results traceable to the SI.
	4.17 lbf·ft to 16.66 lbf·ft	0.06 lbf·ft	
	16.67 lbf·ft to 50 lbf·ft	0.19 lbf·ft	
	50.01 lbf·ft to 100 lbf·ft	0.38 lbf·ft	
	100.01 lbf·ft to 200 lbf·ft	0.76 lbf·ft	
	200 lbf·ft to 300 lbf·ft	1.2 lbf·ft	
Non Transducer Tools including Mechanical Torque Wrenches and Analog Torque Analyzers ^{FO}	0.5 lbf·ft to 4.16 lbf·ft	0.02 lbf·ft	Master Transducers calibrated with results traceable to the SI.
	4.17 lbf·ft to 16.66 lbf·ft	0.07 lbf·ft	
	16.67 lbf·ft to 50 lbf·ft	0.21 lbf·ft	
	50.01 lbf·ft to 100 lbf·ft	0.41 lbf·ft	
	100.01 lbf·ft to 200 lbf·ft	0.82 lbf·ft	
	200 lbf·ft to 300 lbf·ft	1.3 lbf·ft	
Power Tools without Transducers including Air/Pneumatic ^{FO}	0.5 lbf·ft to 4.16 lbf·ft	0.11 lbf·ft	Master Transducers calibrated with results traceable to the SI.
	4.17 lbf·ft to 16.66 lbf·ft	0.43 lbf·ft	
	16.67 lbf·ft to 50 lbf·ft	1.3 lbf·ft	
	50.01 lbf·ft to 100 lbf·ft	2.6 lbf·ft	
	100.01 lbf·ft to 200 lbf·ft	5.2 lbf·ft	
	200 lbf·ft to 300 lbf·ft	7.8 lbf·ft	
	300.01 lbf·ft to 1 000 lbf·ft	26 lbf·ft	



Certificate of Accreditation: Supplement

ITRC

38130 Airport Parkway, Unit #1, Willoughby, OH 449094
Contact Name: John Frangowlakis Phone: 440-622-9540

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.